

VERIFICATION OF COMPLIANCE

No.: LVD SHES2209017159MI
Applicant: Zhejiang Chint Instrument & Meter Co., Ltd.
Wenzhou Bridge Industrial Zone, Yueqing, Zhejiang, P.R. China
Manufacturer: Same as applicant
Product Name: Smart Power Sensor
Product Description: Smart Power Sensor
Model No.: DTSU666-H, DTSU666-H 250A/50mA
Rating: 3x230/400V, 50/60Hz, 1.5W, CT Transformation ratio:
250A/50mA or 100A/40mA
Protection against Electric Shock: Class II
Additional Information: None
Sufficient samples of the product have been tested and found to be in conformity with
Test Standard: EN 61010-1:2010/A1:2019
as shown in the
Test Report Number(s): SHES220901715901

This Verification of Compliance has been granted to the applicant based on the results of tests, performed by Laboratory of SGS-CSTC Standards Technical Services Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant harmonized standards under the Low Voltage Directive 2014/35/EU. The CE marking as shown below can be affixed, under the responsibility of the manufacturer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives. The affixing of the CE marking presumes in addition that the conditions in annexes III and IV of the Directive are fulfilled.



Andrew Zhai

Technical Manager

SGS-CSTC

This verification is issued by the company under its General Conditions of Services accessible at

<https://www.sgs.com/en/terms-and-conditions>. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this verification. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Member of SGS Group (Société Générale de Surveillance)

2022-09-28